



LAWRENCE
LIVERMORE
NATIONAL
LABORATORY

High-Resolution Spectroscopy of K-shell Praseodymium with a High-Energy Calorimeter

D. B. Thorn, G. V. Brown, J. H. T. Clementson, H.
Chen, M. H. Chen, P. Beiersdorfer, K. R. Boyce, C. A.
Kilbourne, F. S. Porter, R. L. Kelley

July 3, 2007

Canadian Journal of Physics

Disclaimer

This document was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor the University of California nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or the University of California, and shall not be used for advertising or product endorsement purposes.

High-Resolution Spectroscopy of K-shell Praseodymium with a High-Energy Microcalorimeter.

Daniel B. Thorn,^{*} Greg V. Brown, Joel H. T.

Clementson, Hui Chen, Mau Chen, and Peter Beiersdorfer

Lawrence Livermore National Laboratory, Livermore, CA 94550, USA

Kevin R. Boyce, Caroline A. Kilbourne, F. Scott Porter, and Richard L. Kelley

Goddard Space Flight Center, NASA, MD 20771, USA

Abstract

We present a measurement of the K-shell spectrum of He-like through Be-like praseodymium ions trapped in the Livermore SuperEBIT electron beam ion trap using a bismuth absorber pixel on the XRS/EBIT microcalorimeter. This measurement is the first of its kind where the $n=2$ to $n=1$ transitions of the various charge states are spectroscopically resolved. The measured transition energies are compared with theoretical calculations from several atomic codes.

^{*}also at Department of Applied Science, University of California, Davis, CA 95616, USA

I. INTRODUCTION

The K-shell spectra of highly ionized few-electron systems have been of great interest for many years. Early measurements focused on relatively low- Z elements such as sulfur, chlorine and argon [1–3] to record the K-shell spectra of helium-like and hydrogen-like ions. This interest in K-shell spectra stems from the desire to measure QED and relativistic effects in the strong field of a heavy nucleus. The reason is that QED corrections and relativistic effects have steep scaling laws that increase their magnitude strongly with increasing atomic number Z . As a result, precise spectroscopic measurements on high- Z systems provide a valuable probe of strong field QED and higher order relativistic effects.

There are, however, two difficulties in obtaining K-shell spectra of heavy highly charged ions. The first, is that high- Z highly charged ions are hard to produce in sufficient quantities. The second, is that measuring the energy of the resulting x-ray radiation to high enough precision required to test QED is limited by the availability of suitable detectors.

The early lower- Z experiments, like those with argon and chlorine, used accelerators to produce the ions and crystal spectrometers to record spectra with high energy resolution. Subsequently, higher- Z ions, i.e. krypton and xenon, were investigated. The krypton measurements were performed with crystal spectrometers as in the lower- Z experiments [4, 5]. Krypton, however, is about the highest- Z system where traditional crystal spectroscopy can be used. This is because the reflectivity for higher-energy photons becomes too low to allow measurements of sufficient statistical quality given the available photon flux.

Conversely, the xenon measurements took a different approach from the lower- Z ions. The earlier measurement by Briand et al. [7], used a high purity germanium detector on the GANIL heavy-ion accelerator to measure both the helium-like and hydrogen-like spectra. But the low intrinsic resolving power of the germanium detector did not allow for as high a precision result as the crystal spectrometers did for the lower- Z ion measurements. The later measurement by Widmann et al. [6], tried to remedy this problem by using a hybrid instrument of a transmission crystal spectrometer and a germanium detector on the SuperEBIT electron beam ion trap and obtained results of higher precision. However, the hybrid method suffered from a low count rate and was extremely time consuming. Thus for measurements of the spectra of ions with higher- Z than xenon, only germanium detectors with their inherently low resolving power for atomic transitions have so far been employed

[8].

To make progress in measuring the K-shell spectra from high-Z highly charged ions, detectors with high-resolving power and high-quantum efficiency are needed. Ultra-low temperature detectors that work by measuring the heat associated with a photon absorbed in a material, so called microcalorimeters, promise to meet both of these criteria. The absorbing material, in principle, can be made of a high-Z element or compound thick enough to stop even the most energetic photons with a reasonable quantum efficiency. In addition, these detectors have already shown the ability to accurately measure the energy of x-ray and γ -ray photons at energies ≤ 60 keV [9, 10].

In this paper we present the first measurement of its kind where the $n=2$ to $n=1$ transitions of few-electron highly ionized ions with $Z \geq 54$ are resolved. Using a microcalorimeter we were able to measure the transition energies of the resonance and forbidden lines in helium-like through beryllium-like praseodymium.

II. METHOD AND RESULTS

The ions were produced at the Livermore SuperEBIT electron beam ion trap facility. Neutral praseodymium was injected into the trap using a laser ablation injection system [11]. The electron beam current was varied from 140-160 mA and the electron beam energy was set at 116 keV. The beam and trapping conditions created a charge balance of helium-like through beryllium-like praseodymium with a small amount of the hydrogen-like charge state.

Two detectors were used in the measurement. The first was a high purity germanium detector. It was used to monitor trapping conditions as its large bandwidth makes it possible to monitor the radiative recombination (RR) part of the spectrum, which gives information on the relative abundances of the charge states in the trap, while simultaneously measuring the direct excitation (DE) part of the spectrum [12]. The second detector was a high-energy microcalorimeter.

The high-energy microcalorimeter used in this measurement was originally designed for x-ray astrophysics and used $8\ \mu\text{m}$ HgTe absorbers attached to silicon thermistors. This choice of absorber material and thickness allowed for the best quantum efficiency (QE) through 7 keV [13–15], while achieving an energy resolution of 6 eV. To achieve a higher QE for

photons above 30 keV the microcalorimeter used in this experiment [10], while of the same design as the one used for astrophysics, differs in that four of the HgTe pixels were replaced by bismuth absorbers with a thickness of 30 μm . This change in the absorber material and thickness did affect the energy resolution obtained. However, the achieved FWHM of one of the pixels was 80 eV at energies near 37 keV giving a resolving power, $E/\Delta E$, of over 450. This is a vast improvement over the Ge detector's resolving power of approximately 80. A second pixel failed to operate properly and the energy resolution of the remaining two pixels was about 150 eV.

Knowing the quantum efficiency of the detector is important when trying to extract cross-sections and other features dependent on the number of events that take place. Because of this, a measurement of the QE of the bismuth pixels were made. The measurement was made relative to the germanium detector which has a QE of 100 percent at 37 keV. The measured QE of the high-energy pixel at a photon energy of 37 keV is close to 40 % as shown in figure 1. This agreed with the theoretical calculation based on the thickness of the absorbing mass of bismuth. For comparison to theory we used the Mass-Energy Absorption Coefficient [16] which takes into account other processes by which x-ray radiation energy is lost in a material.

The data from the pixel with the highest resolution was collected over a period of 40 hours as shown in figure 2. There was no external calibration (such as a radioactive source) for the calorimeter so the strongest lines were used as a relative calibration for the other lines. We took the theoretically accepted values of Drake's calculations for the He-like system [17] and used the resonance line, w, at 37002.7 eV, and the forbidden line, z, at 36292.4 eV, as energy references. A linear fit was used to establish an energy scale for the measurement. A test of the calibration is provided by determining the energy of the intercombination line, y. The intercombination line, according to Drake, has an energy of 36390.4 eV. The value determined by our procedure of 36389.1 ± 6.8 eV agrees with the theoretical value within the limits of the experimental uncertainty.

The resolving power of the high-energy microcalorimeter pixel can be appreciated when compared with the spectrum taken with the germanium detector as shown in figure 3. In the germanium detector spectrum there are only 3 features visible. The first is associated with Lyman- α_1 . The other two features are associated with $2p_{3/2} \rightarrow 1s_{1/2}$ and $2s_{1/2}, 2p_{1/2} \rightarrow 1s_{1/2}$ type transitions in helium-like through beryllium-like praseodymium. In the calorimeter

spectrum the individual transitions of each charge state can clearly be seen.

This increased spectral resolution provided by the calorimeter allows for a more stringent test of theoretical transition energies of the various charge states. Calculations for the helium-like through beryllium-like charge states were made by using the general relativistic atomic structure package (GRASP) [18], and the flexible atomic code (FAC) [19]. Table 1 shows the experimental values determined for the various spectral features relative to the calibration lines. As expected the best agreement with theory is seen with the more highly charged helium-like and lithium-like species. The main source of error in the determination of the measured energies is due to statistics.

III. CONCLUSION

The spectrum obtained with the single bismuth pixel demonstrates that microcalorimeters are highly promising for high-Z K-shell spectroscopic measurements. A relative measurement of 6.8 eV was obtained for the intercombination line, y, in helium-like praseodymium while an 8.5 eV measurement was obtained for the lithium-like lines and a measurement of 31 eV was made for the beryllium-like system. Such a measurement would not have been possible using standard germanium detector technology. It is even more striking when considering that the pixel was not optimized for the calorimeter and simply represents a first attempt. Thus, in the future these devices working on a similar type of experiment can be expected to have far superior resolving powers approaching or exceeding a $E/\Delta E$ of 1000. Furthermore, by making arrays of such pixels it is possible to obtain a very high statistical certainty on the measurement in question.

Acknowledgments

We would like to thank John Gyga and Regis Brekosky at NASA/GSFC for their part in the design and fabrication of the high-energy microcalorimeter and Ed Magee at LLNL for his help in the operation and maintenance of SuperEBIT. This work was performed under the auspices of the U.S. Department Of Energy by the University of California Lawrence

- [1] J. P. Briand, J. P. Moss, P. Indelicato, et. al., *Phys. Rev. A* **28**, 1413 (1983).
- [2] L. Schleinkofer, F. Bell, H. -D. Betz, et. al., *Phys. Scripta* **25**, 917 (1983).
- [3] R. D. Deslattes, R. Schuch, and E. Justiniano, et. al., *Phys. Rev. A* **32**, 1911 (1985).
- [4] J. P. Briand, M. Tavernier, P. Indelicato, et. al., *Z. Phys. A* **318**, 1 (1984).
- [5] K. Widmann, P. Beiersdorfer, V. Decaux, et. al., *Phys. Rev. A* **53**, 2200 (1996).
- [6] K. Widmann, P. Beiersdorfer, G. V. Brown, et a., *AIP Conf. Proc.* **506**, 444 (2000).
- [7] J. P. Briand, P. Indelicato, A. Simionovici, et. al., *Europhys. Lett.* **9**, 225 (1989).
- [8] J. P. Briand, P. Chevallier, P. Indelicato, et. al., *Phys. Rev. Lett.* **65**, 2761 (1990).
- [9] B. R. Beck, J. A. Becker, P. Beiersdorfer, et. al., *Phys. Rev. Lett.* **98**, 142501 (2007).
- [10] F. S. Porter, G. V. Brown, K. R. Boyce, et. al., *Rev. Sci. Instrum.* **75**, 3772 (2004).
- [11] A. M. Niles, E. W. Magee, D. B. Thorn, et. al., *Rev. Sci. Instrum.* **77**, 10F106 (2006).
- [12] J. R. Crespo López-Urrutia, P. Beiersdorfer, D. W. Savin, et. al., *Phys. Rev. Lett.* **77**, 826 (1996).
- [13] R. L. Kelley, S. H. Moseley, C. K. Stahle, et. al., *Journal of Low Temperature Physics* **93** 225 (1993).
- [14] C. K. Stahle, C. A. Allen, K. R. Boyce, et. al., *Nucl. Instrum. Methods A* **520**, 466 (2004).
- [15] R. L. Kelley, K. Mitsuda, C. A. Allen, et. al., *Astron. Soc. Japan* **59**, 77 (2007).
- [16] S. M. Seltzer, *Rad. Res.* **136**, 147 (1993).
- [17] G. W. F. Drake, *Can. J. Phys.* **66**, 586 (1988).
- [18] I. P. Grant, B. J. McKenzie, P. H. Norrington, et al., *Comput. Phys. Commun.* **21**, 207 (1980).
- [19] M. F. Gu, submitted to *Can. J. Phys.* (2007).




FIG. 1: Quantum efficiency of $30\ \mu\text{m}$ bismuth.




FIG. 2: The high resolution spectrum of He-like through Be-like praseodymium using the high-energy microcalorimeter




FIG. 3: Spectrum of K-shell praseodymium obtained with the germanium detector.

TABLE I: Spectral Lines from XRS K-shell Praseodymium Spectrum

Label	Transition	Energy (eV)			
		Measurement	Theory		
			a	b	c
w ^d	$1s2p\ ^1P_1 \rightarrow 1s^2\ ^1S_0$	-	-	-	37002.7
q	$1s2s2p\ ^2P_{3/2} \rightarrow 1s^22s\ ^1S_{1/2}$	36886.8 ± 8.5	36880.35	36877.39	-
β	$1s2s^22p\ ^1P_1 \rightarrow 1s^22s^2\ ^1S_0$	36775 ± 31	36802.71	36798.39	-
y	$1s2p\ ^3P_1 \rightarrow 1s^2\ ^1S_0$	36389.1 ± 6.8	-	-	36390.4
z ^d	$1s2p\ ^3S_1 \rightarrow 1s^2\ ^1S_0$	-	-	-	36292.4
p	$1s2s^2\ ^2S_{1/2} \rightarrow 1s^22p\ ^2P_{1/2}$	36088.2 ± 8.5	36089.01	36085.80	-

^a Present calculation using GRASP [18]

^b Present calculation using FAC [19]

^c [17]

^d Reference line [17]